



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

05-1200 DODECANE

LOT NO: 5236:7

PURITY: 99 + %

GC Injection : On column
 Detector : FID
 Column : CP-WAX-52 CB, 30m x 0.32mm
 film 0.25um
 Carrier : Helium

Temperature
Column
 Initial : 34°C
 Final : 240°C
 Rate : 3°C/min
Injector : 130°C + 140°C/min. Final 240°C
Detector : 240°C

