



INDOFINE Chemical Company, Inc.

121 Stryker Lane, Bldg. 30, Suite 1 • Hillsborough, NJ 08844 • U.S.A.

Phone: (908) 359-6778 • FAX: (908) 359-1179

website: www.indofinechemical.com

e-mail: chemical@indofinechemical.com

05-1800 OCTADECANE

LOT NO: 9217-6

PURITY: >99%

GC Injector : On column
 Detector : FID
 Column : DB-wax, 30m; 0.32mm
 Carrier : He

Temperature

Column

Initial : 34°C 3 min

Final : 220°C

Rate : 3°C / min

Injector : 130°C

Detector : 250°C

